


Table 5 Rules

- TT-5 is for Bends Pain Only
 - Altitude Chamber Induced, Within 6 hours
 - Dive Induced, Within 2 Hours of Surfacing
- Do Not Use This Treatment For Any Other Disorder
 - Any Neurological Symptoms Require TT-6
- If Any Sensation Persists after 10 Minutes @ 60 FSW
 - Go To TT-6
- If Any Discomfort Recurs during TT-5, Go To TT-6
- Inside Observer Takes 10 min O₂ during 2nd O₂ period at 60 FSW, and 30 min during ascent to 30 FSW
- Monoplace Chambers Must Use BIBS for Air Breaks
- Note: Navy TT-5 Omits the Second Air Break at 60FSW and does the Descent on O₂

Treatment Table 5



Descent Rate = 25 ft/min


Ascent Rate = 1 ft/min

Total Elapsed Time: 140
minutes (not including
descent time)

Table 6 Rules

- TT-6 is for Type I and II DCS
- This Table Is Mandatory For Bends Pain More Than 10 Minutes at 60 FSW; All Neurological DCS; Chokes
- May Extend To A Maximum of 5, 20 min Periods @ 60 FSW
- Minimum of 3, 20 min Periods at 60 FSW
- Extend to One O2 Period Completely ASx, or to Max
- May Extend Up To 2 Additional Hrs of O2 at 30 FSW
- Inside Observer Takes 20 min O2 during 3rd O2 period at 60 FSW, Last 20 min at 30 FSW, and 30 min during ascent to Surface. Any Ext @ 60 FSW - add 30 min during ascent to 30 FSW
- USAF May Use USAF or USN Schedule @ 30 FSW
- Monoplace Chambers Must Use BIBS for Air Breaks

USN Treatment Table 6




Descent Rate = 25 ft/min

Ascent Rate = 1 ft/min

Total Elapsed Time: 285
minutes (not including
descent time)

USAF Treatment Table 6



Descent Rate = 25 ft/min

Ascent Rate = 1 ft/min

Total Elapsed Time: 285
minutes (not including
descent time)

Carbon Monoxide Rules

- Treat All With Any Hx CNS Symptoms (other than mild H/A only) Regardless Of CO-Hgb Level
- May Extend To A Maximum of 6, 25 min O2 Periods @ 33 FSW
- Extend Based on Clinical Assessment
- Inside Observer Takes 25 min O2 during last O2 period at 30 FSW, and 10 min during ascent to Surface. Any Ext @ 30 FSW - add 10 min O2 for each extension.
- Tailing Dives @ 45 FSW may be needed in severe cases
 - 3, 30 min O2 periods w/ 10 min air breaks
- Monoplace Chambers Must Use BIBS for Air Breaks

Carbon Monoxide Treatment Table

